

Xenosys.Co.,Ltd.

DECLARATION OF CONFORMITY

Product submitted..... : Medical Head Lamp, Class 1 Device
Model: L2S15
Rated Input: AC 100/240 V; 50/60 Hz; 0.3 A

Applicant(Manufacturer) ... : Xenosys.Co.,Ltd
2 Venture-ro, 100 beon-gil, Yeonsu-gu, Incheon 22013 Republic of Korea,

Testing Laboratory..... : SGS Korea Co., Ltd. (Gunpo Laboratory)


Test Report Number(s)..... : EMC : F609501/RF-EMG005690
Safety : F690501/RF-SAF007688
Usability: F360501/RF-SAF007703

Order Number(s) : G-44-2015-01631

Specification Requested ... : EMC : EN 60601-1-2:2007
EN550152013
EN61000-3-2:2014
EN61000-3-3:2013
Safety : IEC60601-1
Usability: IEC60601-1-6:2010

Conclusion

Based on a review of the test report, this apparatus meets the requirements of the above standards and hence fulfills the requirements of Directive 93/42/EEC.



A handwritten signature in black ink, appearing to be 'Ha-deuk Lee', is written over a light-colored rectangular stamp or seal.

.....
Ha-deuk Lee / President of Xenosys Co., Ltd.

April 20, 2018